

Sample Name: Cu-50
Sample ID :
Time Processed: 11:42:08
Month-Day-Year Processed :04/14/2026

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
Total Flow:1 ml/min
Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.00	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-SP 4.6 x 250 mm

Equipment:ZJ19010014

<Chromatogram>



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	2.731	20291737	1941366	100.000
Total		20291737	1941366	100.000

The Qualitative Detection of Endotoxin

Sample Name : CU-50

Sample ID :

Date Processed(MM/DD/YYYY) :04/14/2026

(Sample ID final concentration 1 mg/ml)

LAL sensitivity: 0.02 EU/ml

Sample Endotoxin level:

1. The endotoxin level in sample was detected using Horseshoe Crab Reagent, the phenomenon is listed as followed:

Table 1:

Inspection tube	0.1ml LAL+0.1ml LAL Reagent Water+0.1ml Detection solution	Parallel	
		1	2
Sample	0.1ml LAL+0.1ml LAL Reagent Water+0.1ml Sample solution	-	-
Sample positive control	0.1ml LAL+0.1ml LAL Reagent Water+0.1ml 2 λSample positive solution	+	+
Positive control	0.1ml LAL+0.1ml LAL Reagent Water+0.1ml positive solution	+	+
Negative control	0.1ml LAL+0.1ml LAL Reagent Water+0.1ml LAL Reagent Water	-	-


Note: The test result is positive which expressed by "+", and negative is expressed by "-".

2. Result:

The endotoxin level in sample is: <0.02EU/mg

MS Spectrum

Positive
Intensity



Date Acquired : 2026/4/14 12:46:55
Sample Name :
Sample ID : Cu-50